

Question:

Please advise dimensions of the biosafety cabinet, including technical specifications, specifications of ULPA filter and outer surface layer offers antimicrobial characteristics.

Answer:

The biosafety cabinet is intended for single user and based on available working space of the laboratory, so the suggested dimension is: External width: < 1.2m, external depth: < 800mm, external height: < 1600 mm.

For the other technical specifications, including ULPA filter, please refer to the respective details in the posted RFQ.